

My No: MSD/SCU(S)/Cir/2025,
Medical Supplies Division,
357, Ven. Baddegama Wimalawansa Thero Mawatha,
Colombo 10.
Tel: (011)2694114/13
02-09-2026

Circular No: **MSD/Q/S/2026/10**

To :

DDG/NHSL – Colombo,
DDG/NH – Kandy,
DDG/NH – Galle,
All Provincial Directors of Health Services,
All Regional Directors of Health Services,
All Directors of Teaching Hospitals,
All Directors of Specialized Campaigns,
Director, N.I.H.S Kalutara,
All MOIC/MOH- Institutions,
MOIC/Medical Institutions of Line Ministry.

REPORT ON FAILING SAMPLE – Colostomy Appliance Set Complete, comprising; 02 Nos. base plates 50mm diameter (approximately) with support flange, adjustable belt and 10 Nos. drainable bags with 2 Nos. clamp

MSD Order List No : **2025/SPC/N/R/S/00034**

PA Order No: **DHS/M/SA/WW/01/020DC/25**

Re-above, details of the letter NMRA/MD/QF/02/2025 dated 25th August 2026 received from CEO/NMRA are circulated herewith for your information and necessary action, please

(NMRA correspondence was received by MSD on 02.09.2026, and the SWASTHA System notification was shared on the same day)

1.1) Name of the product/Item : **Colostomy Appliance Set Complete, comprising; 02 Nos. base plates 50mm diameter (approximately) with support flange, adjustable belt and 10 Nos. drainable bags with 2 Nos. clamp**

Colostomy Appliance Set Complete (BAO-Health)

SR : **12200203**

1.2) Manufacturer : **BAO Health Medical Instruments Co., Ltd, 2nd Floor of 8-2 Building, Gaoqiao Industrial Zone, Tongxiang City, Zhejiang, 314515, P.R. China**

1.3) Batch Details :

Batch No	Date of Manufacture	Date of Expiry
DD202601-0101	2026 February	2029 February

1.4) Reference of NMRA : **To the decision of the 107th Medical Device Evaluation Committee (MDEC), which recommended to withhold the above batch**

1.5) Nature of Defect : **Less adhesive and frequently leaking**

2. CEO-NMRA Instruction : **To withhold the above mentioned batch from use Immediately**

3. Considering above facts, you are requested the following

3.1) **To withhold the above mentioned batch (as detailed 1.1-1.3 above) from use immediately**

4. Please note: It would be your responsibility, in addition to point 3,

4.1) To follow the instructions given by Secretary of Health, under the circular No: 01/2021 dated 08.12.2021.

5. 5.1) Please bring the contents of this circular to all concerned in your Province/ Region/ Institution.

5.2) Please Note: It is the responsibility of all RDHS to ensure that the copies of this circular is circulated among all the heads of the institutions under your preview.

6. Previous Circulars

Circular No	Batch	Decision
MSD/Q/S/2025/11	DD202405-01	Withhold
MSD/Q/S/2025/18	DD202311-19	Withhold
MSD/Q/S/2026/5	DD202508-1301	Withhold
MSD/Q/S/2026/8	DD202508-1301	Withdraw
	DD202311-19	
	DD202405-01	


Director,
Medical Supplies Division,

Dr. DEDUNU DIAS
Director - (Acting)
Medical Supplies Division
Rev. Baddegama Wimalawansa Thero Mw.,
Colombo 10.

Cc:

- | | | |
|--|---|--------------------|
| 1. Auditor General - f.i. | 2.DGHS-f.i. | 3. DDG(M/S) - f.i. |
| 4. Chairman/SPC – f.i. & n.a. | | |
| 5.CEO/NMRA – f.i. | 6. DD/MSD - f.i. & n.a | |
| 7. D/NMQAL – f.i. | | |
| 8. Chief Internal Auditor/ MOH-f.i. | 9.D/(Stock Verific.)/MOH - f.i. | |
| 10. Chief Accountant (Finance) / MSD – f.i. | | |
| 11. Chief Accountant (Supply) / MSD – f.i. | | |
| 12. AD/ICT Unit – To publish in MSD web site | | |
| 13. AD(S)/MSD – f.i. & n.a. | | |
| 14. AD (M&C) – f.i. | | |
| 15. AD (Stores)/MSD - f.i. | 16. AD(Dispatch)/MSD – f.i. | |
| 17. Secretary / MDEC- f.i. | | |
| 18. All DP/DPDHS - f.i. & n.a. | 19. SP / MSD - f.i. | |
| 20. HSCO (S) - f.i. & n.a. | | |
| 21. SCO (S/WH 12) – f.i. & n.a. | | |
| 22. SCO (D/WH 12)- f.i & n.a. | 23.MSA-WH (12)/MSD - f.i. & n.a. | |
| 24. MO (Technical unit)/MOH- f.i & n.a. | 25. H/Destruction Unit (MSD) - f.i & n.a. | |